

**Certificate of Analysis  
GM501 Flush**

Batchnumber	FLUSH-2006
Production date	06 August 2020
Expiry date	28 February 2021
Total volume (mL)	61900

	<b>Results</b>	<b>Specifications</b>
pH	7,37	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 21 August 2020

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*