

**Certificate of Analysis  
GM501 Flush**

Batchnumber	FLUSH-2101
Production date	21 January 2021
Expiry date	31 July 2021
Total volume (mL)	60400

	<b>Results</b>	<b>Specifications</b>
pH	7,34	7,20-7,40
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 09 February 2021

Released by Vanessa Bierlich

Signature