

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-2106
Production date	14 May 2021
Expiry date	30 November 2021
Total volume (mL)	24000

	Results	Specifications
pH	7,38	7,20-7,40
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 31 May 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann