

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-2109
Production date	27 August 2021
Expiry date	28 February 2022
Total volume (mL)	59700

	Results	Specifications
pH	7,28	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 13 September 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann