

Certificate of Analysis GM501 Flush

Batchnumber	FLUSH-2203
Production date	25 February 2022
Expiry date	31 August 2022
Total volume (mL)	63000

	Results	Specifications
pH	7,32	7,20-7,40
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 16 March 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann