

Certificate of Analysis GM501 Flush

Batchnumber	FLUSH-2205
Production date	06 May 2022
Expiry date	30 November 2022
Total volume (mL)	73200

	Results	Specifications
pH	7,36	7,20-7,40
Osmolality	281	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 23 May 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann