

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-2207
Production date	02 August 2022
Expiry date	28 February 2023
Total volume (mL)	73400

	Results	Specifications
pH	7,37	7,20-7,40
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 17 August 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann