

**Certificate of Analysis
GM501 Flush**

Batchnumber FLUSH-2402
 Production date 01 February 2024
 Expiry date 31 August 2024
 Total volume (ml) 42300

Parameter	Results	Specifications
pH	7,38	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse embryo test (1-cell)	100	≥ 80 %

Release date 16 February 2024

Released by Nicole Geiger

Signature

