


**Certificate of Analysis
GM501 Flush**

Batchnumber FLUSH-2403
 Production date 01 March 2024
 Expiry date 30 September 2024
 Total volume (ml) 72500

Parameter	Results	Specifications
pH	7,21	7,20-7,40
Osmolality	281	270-290 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse embryo test (1-cell)	85,71	≥ 80 %

Release date 19 March 2024
 Released by Nicole Geiger
 Signature



REV 07_00 / 2023-09-15