

Certificate of Analysis
GM501 Flush

Batchnumber FLUSH-2504

Production date 02 May 2025

Expiry date 30 November 2025

Total volume (ml) 115.500

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,35	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Sterility test (performed according to the current PH. Eur. 2.6.1 / USP <71>)	No growth	No growth
Concentration heparin	3,1	2-3,75 IU/mL
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse embryo test (1-cell)	100	≥ 80 %

Release date 20 May 2025

Released by Nicole Geiger

Signature

pp. N. Geiger

REV 08_00 / 2024-08-16