

**Certificate of Analysis
GM501 Flush**

Batchnumber FLUSH-2509

Production date 07 November 2025

Expiry date 31 May 2026

Total volume (ml) 109800

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,31	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Sterility test (performed according to the current PH. Eur. 2.6.1 / USP <71>)	No growth	No growth
Concentration heparin	2,38	2-3,75 IU/mL
Endotoxin	< 0,06	< 0,25 EU/mL
Mouse embryo test (1-cell)	90	≥ 80 %

Release date 24 November 2025

Released by Joana Sturm

Signature



REV 08_00 / 2024-08-16

