

**Certificate of Analysis
GM501 Flush**

Batchnumber FLUSH-2604
 Production date 12 March 2026
 Expiry date 30 September 2026
 Total volume (ml) 116100

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,28	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Sterility test (performed according to the current PH. Eur. 2.6.1 / USP <71>)	No growth	No growth
Concentration heparin	2,41	2-3,75 IU/mL
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse embryo test (1-cell)	100	≥ 80 %

Release date 27 March 2026

Released by Dr. René Richarz

Signature



REV 08_00 / 2024-08-16