

Certificate of Analysis GM501 Flush

Batchnumber	FLUSH-2610
Production date	10 September 2026
Expiry date	31 March 2027
Total volume (ml)	111000
Product was sterilized by aseptic processing techniques	

Parameter	Results	Specifications
pH	7,30	7,20-7,40
Osmolality	283	270-290 mOsm/kg
Sterility test (performed according to the current PH. Eur. 2.6.1 / USP <71>)	No growth	No growth
Concentration heparin	2,65	2-3,75 IU/mL
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse embryo test (1-cell)	95	≥ 80 %

Release date	26 September 2026
Released by	Nicole Geiger
Signature	

pp. *N. Geiger*

REV 08_00 / 2024-08-16