# Certificate of Analysis
# GM501 Flush

| | |
|---|---|
| Batchnumber | FLUSH-2611 |
| Production date | 13 October 2026 |
| Expiry date | 30 April 2027 |
| Total volume (ml) | 87700 |

Product was sterilized by aseptic processing techniques

| Parameter | Results | Specifications |
|---|---|---|
| pH | 7,30 | 7,20-7,40 |
| Osmolality | 284 | 270-290 mOsm/kg |
| Sterility test (performed according to the current PH. Eur. 2.6.1 / USP <71>) | No growth | No growth |
| Concentration heparin | 2,53 | 2-3,75 IU/mL |
| Endotoxin | < 0,05 | < 0,25 EU/mL |
| Mouse embryo test (1-cell) | 100 | ≥ 80 % |

| | |
|---|---|
| Release date | 28 October 2026 |
| Released by | Dr. René Richarz |
| Signature |  |



Gynemed GmbH Co. KG, Wagrienring 24b, 23730 Sierksdorf, Germany
Phone: +49-4561-5132-500, fax: +49-4561-5132-519, e-mail: info@gynemed.de