Certificate of Analysis
GM501 Flush

Batchnumber	FLUSH-2612
Production date	20 October 2026
Expiry date	30 April 2027
Total volume (ml)	38100

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,28	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Sterility test (performed according to the current PH. Eur. 2.6.1 / USP <71>)	No growth	No growth
Concentration heparin	2,83	2-3,75 IU/mL
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse embryo test (1-cell)	90	≥ 80 %

Release date	05 November 2026
Released by	Nicole Geiger
Signature	pp. N. Geiger

Gynemed GmbH Co. KG, Wagrienring 24b, 23730 Sierksdorf, Germany
Phone: +49-4561-5132-500, fax: +49-4561-5132-519, e-mail: info@gynemed.de