

**Certificate of Analysis
GM501 Gradient 45%**

Batchnumber	GRAD45-1938
Production date	19 November 2019
Expiry date	31 May 2021
Total volume (mL)	12250

	Results	Specifications
pH	7,4	7,20-7,60
Osmolality	328	310-340 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,50 EU/mL
Sperm survival Test (4 hours exposure)	98	≥ 80 %
Sperm survival Test (24 hours exposure)	94	≥ 75 %

Release date	04 December 2019
Released by	Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann