

**Certificate of Analysis  
GM501 Gradient 45%**

Batchnumber GRAD45-2425  
 Production date 28 October 2024  
 Expiry date 30 April 2026  
 Total volume (ml) 17950

<b>Parameter</b>	<b>Results</b>	<b>Specifications</b>
pH	7,3	7,20-7,60
Osmolality	327	310-340 mOsm/kg
Endotoxin level	< 0,15	< 0,50 EU/mL
Density	1,0582	1,1050-1,1150 g/ml
Sterility test (according to the current Ph. Eur. 2.6.1 / <71>)	No growth	No growth
HSSA (Human sperm survival assay) (4 hours exposure)	98	≥ 80 %
HSSA (Human sperm survival assay) (24 hours exposure)	96	≥ 75 %

Release date 12 November 2024  
 Released by Joana Sturm

Signature



REV 09\_00 / 2024-09-11