

**Certificate of Analysis
GM501 Gradient 45%**

Batchnumber	GRAD45-2428
Production date	29 November 2024
Expiry date	31 May 2026
Total volume (ml)	8850

Parameter	Results	Specifications
pH	7,33	7,20-7,60
Osmolality	334	310-340 mOsm/kg
Endotoxin level	< 0,15	< 0,50 EU/mL
Density	1,0574	1,0500-1,0700 g/ml
Sterility test (according to the current Ph. Eur. 2.6.1 / <71>)	No growth	No growth
HSSA (Human sperm survival assay) (4 hours exposure)	97	≥ 80 %
HSSA (Human sperm survival assay) (24 hours exposure)	90	≥ 75 %

Release date	16 November 2024
Released by	Joana Sturm
Signature	



REV 09_00 / 2024-09-11