

**Certificate of Analysis
GM501 Gradient 45%**

Batchnumber	GRAD45-2517
Production date	13 June 2025
Expiry date	31 December 2026
Total volume (ml)	8600

Parameter	Results	Specifications
pH	7,31	7,20-7,60
Osmolality	321	310-340 mOsm/kg
Endotoxin level	< 0,15	< 0,50 EU/mL
Density	1,0590	1,0500-1,0700 g/ml
Sterility test (according to the current Ph. Eur. 2.6.1 / <71>)	No growth	No growth
HSSA (Human sperm survival assay) (4 hours exposure)	100	≥ 80 %
HSSA (Human sperm survival assay) (24 hours exposure)	100	≥ 75 %

Release date	01 July 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 09_00 / 2024-09-11