

**Certificate of Analysis  
GM501 Gradient 90%**

Batchnumber	GRAD90-2228
Production date	19 August 2022
Expiry date	28 February 2024
Total volume (mL)	6200

	<b>Results</b>	<b>Specifications</b>
pH	7,34	7,20-7,60
Osmolality	339	320-350 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,50 EU/mL
Density	1,1106	1.1050-1.1150 g/mL
Viscosity in cP (26°C)	1,64	< 1,65 cP
Sperm survival Test (4 hours exposure)	100	≥ 80 %
Sperm survival Test (24 hours exposure)	87	≥ 75 %

Release date 05 September 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*