

**Certificate of Analysis
GM501 Gradient 90%**

Batchnumber	GRAD90-2428
Production date	29 November 2024
Expiry date	31 May 2026
Total volume (ml)	12450

Parameter	Results	Specifications
pH	7,32	7,20-7,60
Osmolality	343	320-350 mOsm/kg
Density	1,1103	1,1050-1,1150 g/ml
Endotoxin level	< 0,15	< 0,50 EU/ml
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (Human sperm survival assay) (4 hours exposure)	95	≥ 80 %
HSSA (Human sperm survival assay) (24 hours exposure)	91	≥ 75 %

Release date	16 December 2024
Released by	Joana Sturm
Signature	



REV 09_00 / 2024-09-11