

**Certificate of Analysis
GM501 Gradient 90%**

Batchnumber	GRAD90-2507
Production date	05 March 2025
Expiry date	30 September 2026
Total volume (ml)	19750

Parameter	Results	Specifications
pH	7,31	7,20-7,60
Osmolality	343	320-350 mOsm/kg
Density	1,1086	1,1050-1,1150 g/ml
Endotoxin level	< 0,15	< 0,50 EU/ml
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (Human sperm survival assay) (4 hours exposure)	99	≥ 80 %
HSSA (Human sperm survival assay) (24 hours exposure)	100	≥ 75 %

Release date	21 March 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger