

**Certificate of Analysis
GM501 Gradient 100%**

Batchnumber	GSS-2012
Production date	26 June 2020
Expiry date	31 December 2021
Total volume (mL)	85000

	Results	Specifications
pH	7,3	7,20-7,60
Osmolality	313	300-330 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,50 EU/mL
Density	1,1197	1,1150-1,1250 g/mL
Viscosity in cP (26°C)	1,73	< 1,75 cP
Sperm survival Test (4 hours exposure)	98	≥ 80 %
Sperm survival Test (24 hours exposure)	93	≥ 75 %

Release date 16 July 2020

Released by Vanessa Bierlich

Signature