

**Certificate of Analysis
GM501 Gradient 100%**

Batchnumber GSS-2520
Production date 28 May 2025
Expiry date 30 November 2026
Total volume (ml) 100.000

Product is sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,29	7,20-7,60
Osmolality	313	300-330 mOsm/kg
Endotoxin level	< 0,15	< 0,50 EU/ml
Density	1,1216	1,1150-1,1250 g/ml
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to test medium)	98	≥ 80 %
HSSA (% motility compared with control after 24h exposure to test medium)	90	≥ 75 %

Release date 13 June 2025
Released by Nicole Geiger
Signature

pp. N. Geiger

REV 08_00 / 2025-01-10