
GYNEMED

Certificate of Analysis
GM501 HSA

Batchnumber	HSA-2606
Production date	14 October 2026
Expiry date	31 October 2027
Total volume (ml)	9000
Batchnumber HSA	KP51A6872

Parameter	Results	Specifications
pH	7,33	7,20-7,40
Osmolality	280	265-285 mOsm/kg
Endotoxin level	< 0,2	< 1,0 EU/mL
Concentration HSA	94,7	80-120 g/l
Color	Amber	AMBER
Sterility (fungal and bacterial sterility test)	No growth	NO GROWTH
Clarity	Clear, no particles	Clear, no particles
Mouse Embryo Assay (1-cell to expanded blastocyst within 96 hours)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	11 November 2026
Released by	Nicole Geiger
Signature	

Gynemed GmbH Co. KG, Wagrienring 24b, 23730 Sierksdorf, Germany
Phone: +49-4561-5132-500, fax: +49-4561-5132-519, e-mail: info@gynemed.de