

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-1705
Production date	17 March 2017
Expiry date	31 March 2018
Total volume (kits)	150
Batchnumber HSA	A4DF500171

	Results	Specifications
pH	7,44	7,20-7,60
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,117	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	03 April 2017
--------------	---------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann