

**Certificate of Analysis  
GM501 Hyaluronidase**

Batchnumber	HYA-1709
Production date	31 May 2017
Expiry date	31 May 2018
Total volume (kits)	229
Batchnumber HSA	A4DF500171

	<b>Results</b>	<b>Specifications</b>
pH	7,46	7,20-7,60
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,178	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 15 June 2017

Released by Vanessa Bierlich

Signature