

**Certificate of Analysis  
GM501 Hyaluronidase**

Batchnumber	HYA-1908
Production date	27 March 2019
Expiry date	31 March 2020
Total volume (kits)	351
Batchnumber HSA	B3ALC00323

	<b>Results</b>	<b>Specifications</b>
pH	7,41	7,30-7,60
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,099	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	80,95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 11 April 2019

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*