

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-1912
Production date	16.15.2019
Expiry date	31 May 2019
Total volume (kits)	407
Batchnumber HSA	B3ALC00323

	Results	Specifications
pH	7,46	7,30-7,60
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,114	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 31 May 2019

Released by Dipl.- Ing. Benthe Brauer

Signature