

## Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-1914
Production date	04 June 2019
Expiry date	30 June 2020
Total volume (kits)	370
Batchnumber HSA	B3ALC00323

	Results	Specifications
pH	7,52	7,30-7,60
Osmolality	281	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,159	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	19 June 2019
--------------	--------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

*p.p. Julia Heinzmann*