

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-1922
Production date	26 September 2019
Expiry date	30 September 2020
Total volume (ml)	4330
	B3ALC00323
Batchnumber HSA	K824B6871

	Results	Specifications
pH	7,49	7,30-7,60
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,254	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	11 October 2019
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann