

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-2014
Production date	04 September 2020
Expiry date	30 September 2021
Total volume (ml)	27000
Batchnumber HSA	B3ALD01343

	Results	Specifications
pH	7,52	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,153	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	21 September 2020
--------------	-------------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann