

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-2021
Production date	03 December 2020
Expiry date	31 December 2021
Total volume (ml)	3000
Batchnumber HSA	B3ALE00033

	Results	Specifications
pH	7,44	7,30-7,60
Osmolality	277	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,162	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	18 December 2020
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann