

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-2106
Production date	17 March 2021
Expiry date	31 March 2022
Total volume (ml)	4300
Batchnumber HSA	B3ALE00033

	Results	Specifications
pH	7,47	7,30-7,60
Osmolality	286	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,276	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 06 April 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann