

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-2203
Production date	26 January 2022
Expiry date	31 January 2023
Total volume (ml)	102000
Batchnumber HSA	K052B6875

	Results	Specifications
pH	7,5	7,30-7,60
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,261	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	90	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 14 February 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann