

**Certificate of Analysis  
GM501 Hyaluronidase**

Batchnumber	HYA-2209
Production date	22 April 2022
Expiry date	30 April 2023
Total volume (ml)	3200
Batchnumber HSA	B3ALF01143
Batchnumber HYA	S1026H40

	<b>Results</b>	<b>Specifications</b>
pH	7,52	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,142	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 09 May 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*

REV 08\_00 /2022-05-09