

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-2301
Production date	02 January 2023
Expiry date	31 January 2024
Total volume (ml)	4400
Batchnumber HSA	K222B6875
Batchnumber HYA	S1026H40

	Results	Specifications
pH	7,46	7,30-7,60
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,2	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	01 February 2023
Released by	Nicole Geiger

Signature

