

**Certificate of Analysis  
GM501 Hyaluronidase**

Batchnumber	HYA-2324
Production date	24 August 2023
Expiry date	31 August 2024
Total volume (ml)	3600
Batchnumber HSA	B03H003853
Batchnumber HYA	S1047H40

	<b>Results</b>	<b>Specifications</b>
pH	7,56	7,30-7,60
Osmolality	281	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,08	< 1,00 EU/mL
Mouse Embryo Assay (1-cell)	85,71	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 13 September 2023

Released by Nicole Geiger

Signature

pp. N. Geiger