

**Certificate of Analysis**  
**GM501 Hyaluronidase**

Batchnumber	HYA-2328
Production date	13 October 2023
Expiry date	31 October 2024
Total volume (ml)	3600
Batchnumber HSA	B03H003853
Batchnumber HSA	B03H025323
Batchnumber HYA	S1047H40

Parameter	Results	Specifications
pH	7,52	7,30-7,60
Osmolality	277	270-290 mOsm/kg
Endotoxin	< 0,11	<0,5 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Enzyme activity	67-98	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 30 October 2023

Released by Nicole Geiger

Signature

pp. N. Geiger

REV 09\_00 /2023-09-15