

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber HYA-2412
 Production date 26 March 2024
 Expiry date 31 March 2025
 Total volume (ml) 3600
 Batchnumber HSA B03H025323
 Batchnumber HYA S1047H40

Parameter	Results	Specifications
pH	7,48	7,30-7,60
Osmolality	281	270-290 mOsm/kg
Endotoxin	< 0,13	<0,5 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Enzyme activity	80-118	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 17 April 2024
 Released by Vanessa Bierlich
 Signature

REV 09_00 /2023-09-15