

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-2415
Production date	29 April 2024
Expiry date	30 April 2025
Total volume (ml)	3400
Batchnumber HSA	B03H003853
Batchnumber HYA	S1047H40

Parameter	Results	Specifications
pH	7,46	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Endotoxin	< 0,13	<0,5 EU/ml
Concentration HSA	3,8	3,2-4,8 g/l
Enzyme activity	88-129	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	NO GROWTH	No growth
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	16 May 2024
Released by	Nicole Geiger
Signature	

REV 09_00 /2023-09-15