


**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber HYA-2417
 Production date 15 May 2024
 Expiry date 31 May 2025
 Total volume (ml) 5700
 Batchnumber HSA B03H025323, B03H100413,
 B03H003853
 Batchnumber HYA S1047H40

Parameter	Results	Specifications
pH	7,47	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Endotoxin	< 0,1	<0,5 EU/ml
Concentration HSA	3,8	3,2-4,8 g/l
Enzyme activity	80-118	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 30 May 2024
 Released by Nicole Geiger
 Signature



REV 09_00 /2023-09-15