

**Certificate of Analysis  
GM501 Hyaluronidase**

Batchnumber	HYA-2419
Production date	27 May 2024
Expiry date	31 May 2025
Total volume (ml)	3600
Batchnumber HSA	B03H00385 B03H100413
Batchnumber HYA	S1047H40

<b>Parameter</b>	<b>Results</b>	<b>Specifications</b>
pH	7,44	7,30-7,60
Osmolality	277	270-290 mOsm/kg
Endotoxin	< 0,11	<0,5 EU/ml
Concentration HSA	4,1	3,2-4,8 g/l
Enzyme activity	89-131	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	12 June 2024
Released by	Nicole Geiger
Signature	