

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber HYA-2503
 Production date 21 January 2025
 Expiry date 31 January 2026
 Total volume (ml) 3300
 Batchnumber HSA B03H108433
 Batchnumber HYA S1047H40

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,47	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Endotoxin	< 0,1	<0,5 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Enzyme activity	70-104	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 11 February 2025
 Released by Nicole Geiger
 Signature 