

**Certificate of Analysis
GM501 Hyaluronidase**

| | |
|-------------------|--------------|
| Batchnumber | HYA-2519 |
| Production date | 18 June 2025 |
| Expiry date | 30.06.2026 |
| Total volume (ml) | 5700 |
| Batchnumber HSA | B03J094493 |
| Batchnumber HYA | S2015H40 |

Product was sterilized by aseptic processing techniques

| Parameter | Results | Specifications |
|--|-----------|-----------------|
| pH | 7,45 | 7,30-7,60 |
| Osmolality | 282 | 270-290 mOsm/kg |
| Endotoxin | < 0,07 | <0,5 EU/ml |
| Concentration HSA | 4,0 | 3,2-4,8 g/l |
| Enzyme activity | 43-63 | 40-160 IU/ml |
| Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>) | No growth | No growth |
| Mouse Embryo Assay (1-cell) | 90,48 | ≥ 80 % |

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

| | |
|--------------|---|
| Release date | 07 July 2025 |
| Released by | Nicole Geiger |
| Signature |  |

REV 09_00 /2023-09-15