

**Certificate of Analysis**  
**GM501 Hyaluronidase**

Batchnumber HYA-2540  
 Production date 09 December 2025  
 Expiry date 31 December 2026  
 Total volume (ml) 3600  
 Batchnumber HSA B03J094493  
 Batchnumber HYA S2015H40  
 Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,42	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Endotoxin	< 0,06	<0,5 EU/ml
Concentration HSA	4,0	3,2-4,8 g/l
Enzyme activity	49-71	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 02 January 2026  
 Released by Joana Sturm  
 Signature   
 Signature

REV 09\_00 /2023-09-15