

**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber HYA-2612
 Production date 10 April 2026
 Expiry date 31 March 2027
 Total volume (ml) 3600
 Batchnumber HSA B03J125033
 Batchnumber HYA S2015H40

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,39	7,30-7,60
Osmolality	280	270-290 mOsm/kg
Endotoxin	< 0,05	<0,5 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Enzyme activity	56-83	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 27 April 2026
 Released by Dr. René Richarz
 Signature 

REV 09_00 /2023-09-15