

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-2613
Production date	15 April 2026
Expiry date	31 March 2027
Total volume (ml)	3600
Batchnumber HSA	B03J125033
Batchnumber HYA	S2015H40

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,46	7,30-7,60
Osmolality	281	270-290 mOsm/kg
Endotoxin	< 0,05	<0,5 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Enzyme activity	53-79	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	05 May 2026
Released by	Nicole Geiger
Signature	

REV 09_00 /2023-09-15