

**Certificate of Analysis  
GM501 Hyaluronidase**

Batchnumber HYA-2617  
 Production date 12 May 2026  
 Expiry date 31 May 2027  
 Total volume (ml) 4000  
 Batchnumber HSA B03J125033  
 Batchnumber HYA S2015H40

Product was sterilized by aseptic processing techniques

<b>Parameter</b>	<b>Results</b>	<b>Specifications</b>
pH	7,49	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Endotoxin	< 0,19	<0,5 EU/ml
Concentration HSA	4,0	3,2-4,8 g/l
Enzyme activity	81-119	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 28 May 2026  
 Released by Dr. René Richarz  
 Signature 

REV 09\_00 /2023-09-15