


**Certificate of Analysis
GM501 Hyaluronidase**

Batchnumber	HYA-2621
Production date	08 June 2026
Expiry date	30 June 2027
Total volume (ml)	6000
Batchnumber HSA	B03J125033
Batchnumber HYA	S2015H40

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,47	7,30-7,60
Osmolality	280	270-290 mOsm/kg
Endotoxin	< 0,13	<0,5 EU/ml
Concentration HSA	3,8	3,2-4,8 g/l
Enzyme activity	57-84	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	24 June 2026
Released by	Nicole Geiger
Signature	

REV 09_00 /2023-09-15